

# Method for the determination of lead and its inorganic compounds

---

**German Social Accident Insurance**  
**Deutsche Gesetzliche Unfallversicherung**  
Analytical Subcommittee of the Chemistry Board of Experts\*

<b>Carcinogenic substances</b>	Order number:	BGI 505-73-01
<b>Established methods:</b>	Issued:	January 2009

---

This method has been tested and recommended by the German Social Accident Insurance for the determination of lead and its inorganic compounds at workplaces.

Both personal and stationary sampling can be performed for the assessment of workplaces.

Sampling is carried out with a pump and collection on a membrane filter. Analysis is performed by graphite furnace atomic absorption spectrometry (GFAAS) after acid digestion.

<b>Chemical name:</b>	Lead
<b>CAS number:</b>	7439-92-1
<b>Symbol:</b>	Pb
<b>Atomic mass:</b>	207.19 g/mol

---

\* Please direct correspondence to Berufsgenossenschaft RCI, Fachbereich Gefahrstoffe und biologische Arbeitsstoffe, P.O. Box 101480, 69004 Heidelberg, Germany; analytik@bgrci.de

## Summary

This method can be used with personal or stationary sampling to determine the mean concentration of lead and its inorganic compounds at workplaces over the sampling time.

- Principle:** A pump draws a defined volume of air through a membrane filter. The lead contained in the collected aerosol is determined by means of graphite furnace atomic absorption spectrometry after acid digestion.
- Limit of quantification:** Absolute: 0.044 ng of lead  
Relative: 0.13  $\mu\text{g}/\text{m}^3$  for an air sample of 1.2  $\text{m}^3$ , a sample solution of 20 mL (dilution factor 4) and an injection volume of 20  $\mu\text{L}$ .
- Selectivity:** The selectivity of the method depends above all on the wavelength selected, the absence of spectral interference and the minimization of non-spectral interference. The conditions stated here have proved successful in practice.
- Advantages:** Personal measurements with high sensitivity are possible; the method can be used for short-term measurements.
- Disadvantages:** Time-consuming, requires highly sophisticated apparatus.
- Apparatus:** Sampling system consisting of a pump, volumetric flow meter, filter holder with membrane filter, digestion apparatus, atomic absorption spectrometer with graphite furnace.

## Detailed description of the method

### Contents

1	Equipment, chemicals and solutions
1.1	Equipment
1.2	Chemicals and solutions
2	Sampling procedure
3	Analytical determination
3.1	Sample preparation and analysis
3.2	Operating conditions for atomic absorption spectrometry
4	Evaluation
4.1	Calibration
4.2	Calculation of the analytical result
5	Reliability of the method
5.1	Precision and recovery
5.2	Limit of quantification
5.3	Selectivity
6	Remarks

## 1

### Equipment, chemicals and solutions

#### 1.1

##### Equipment

For sampling:

- Sampling device suitable for collecting the inhalable aerosol fraction GSP 10 sampling head (PGP according to BGIA), e.g. from GSM Gesellschaft für Schadstoffmesstechnik GmbH, 41469 Neuss, Germany
- Pump for sampling with a nominal flow rate of 10 L/min, GSM/SG10, e.g. from Leschke Messtechnik GmbH, 15230 Frankfurt (Oder), Germany
- Membrane filter, diameter 37 mm, pore size 8.0 µm, cellulose nitrate, if possible with a test certificate stating the content of metals, e.g. from Sartorius AG, 37075 Göttingen, Germany
- Volumetric flow meter



Fig. 1 Schematic illustration of a digestion vessel with attached cooler.

For sample preparation:

- Aluminum heating block thermostat with external time/temperature control, operating range up to 200 °C, e.g. from Gebr. Liebisch GmbH & Co, 33649 Bielefeld, Germany
- Graduated digestion vessels with air cooler (see Fig. 1) made of quartz glass (diameter 19 mm, maximum volume 25 mL) with ground connections (NS 19/26), acid-proof 0.2 mL graduation in the range from 15 to 25 mL, e.g. from VWR International GmbH, 40764 Langenfeld, Germany
- Polyethylene stoppers for the digestion vessels (NS 19/26), e.g. from Pöppelmann GmbH & Co, 49378 Lohne, Germany
- Quartz glass rods (diameter approx. 4 mm), fitted with replaceable endpieces of PTFE tube<sup>1)</sup>, e.g. from VWR International GmbH
- 5 L bottle made of perfluoroalkoxy copolymer (PFA) with PTFE dispenser for rinsing the air cooler or preparing dilutions, e.g. Optifix HF dispenser, 30 mL, from Poulten & Graf GmbH (Fortuna), 97877 Wertheim, Germany
- Measuring cylinders made of PFA, 500 mL, 100 mL, 50 mL, e.g. from VIT-LAB GmbH, 64332 Seeheim-Jugenheim, Germany

1) Normal boiling glass capillaries cannot be cleaned without leaving residues.

- Pair of ceramic tweezers to transfer the membrane filters to the digestion vessels, e.g. from PLANO, W. Plannet GmbH, 35578 Wetzlar, Germany

For the analytical determination:

- Atomic absorption spectrometer with equipment for graphite furnace technique and autosampler
- Volumetric flasks made of PFA for standard and calibration solutions with screw caps and ring marks, 500 mL, 100 mL, 50 mL, e.g. from VIT-LAB GmbH
- Disposable polystyrene vessels, volume approx. 1.5 mL for the autosampler, e.g. from Greiner Bio-One GmbH, 72636 Frickenhausen, Germany
- Adjustable piston pipettes with the following volume ranges: 2 to 20  $\mu\text{L}$ , 20 to 100  $\mu\text{L}$ , 30 to 200  $\mu\text{L}$ , 200 to 1000  $\mu\text{L}$ , 1000 to 5000  $\mu\text{L}$ , e.g. Pipetman Ultra, from Gilson International B.V., 65555 Limburg-Offheim, Germany
- Ultrapure water unit, e.g. from Wilhelm Werner GmbH, 51381 Leverkusen, Germany

## 1.2

### Chemicals and solutions

Ultrapure water (specific resistance  $\geq 17 \text{ M}\Omega \times \text{cm}$  at 25 °C)

Nitric acid (65%), with low metal content, e.g. Suprapur, from Merck KGaA, 64271, Darmstadt, Germany

Hydrochloric acid (30%), with low metal content, e.g. Suprapur, from Merck KGaA  
Standard acid digestion mixture according to the DFG [1], prepared from 1400 mL of nitric acid, 570 mL of hydrochloric acid and 130 mL of ultrapure water (equivalent to 2 parts by volume of nitric acid (65%) and 1 part by volume of hydrochloric acid (25%))

Lead Stock Solution 1:  $c = 1000 \mu\text{g/mL}$ , e.g. Product No. 13853 (plasma standard solution) from Alfa Aesar GmbH & Co. KG, 76185 Karlsruhe, Germany

Lead Stock Solution 2:  $c = 25 \mu\text{g/L}$  (25  $\mu\text{L}$  of Stock Solution 1 are placed in a 1000 mL volumetric flask, and the flask is filled to the mark with ultrapure water)

Calibration solutions (see Section 4.1)

Argon 4.6 (purity at least 99.996%)

## 2

### Sampling procedure

A membrane filter is inserted into the filter holder, and then the holder is placed in GSP sampling head, which is connected to the pump. Stationary or personal sampling can be carried out. The flow rate is initially adjusted to 10 L/min. The definition of the inhalable dust is fulfilled by this flow rate [2]. At a sampling time

of two hours this corresponds to an air sample volume of 1.2 m<sup>3</sup>. After sampling, the flow rate must be checked for constancy. If the deviation from the adjusted flow rate is greater than  $\pm 5\%$ , it is advisable to discount the measurement (see BGI 505-0 “General Part”, Section 3).

### 3 Analytical determination

#### 3.1

##### Sample preparation and analysis

The loaded filter is folded using a pair of ceramic tweezers, transferred to the 25 mL digestion vessel and pressed down to the bottom of the vessel by means of a quartz glass rod (see Section 1.1). Then 10 mL of standard acid digestion mixture are added.

An air cooler is attached to the vessel and then the vessel is placed in the aluminum heating block thermostat for two hours under reflux (block temperature: approx. 125 °C).

After cooling down to approx. 50 °C, 10 mL of ultrapure water are carefully added through the air cooler to rinse the cooler and also to dilute the viscous solution.

A brief heating phase follows to achieve homogenization. After cooling, the air cooler and the quartz glass rod are removed, the digestion vessel is sealed with a polyethylene stopper and the volume of the solution (sample solution) is read off. Before analysis, the sample solution must be diluted (see Section 3.2).

A current blank value  $\bar{c}_{Bl(\text{current})}$  is determined for each series of samples. For this purpose a filter that has not been used for sampling is subjected to the entire sample preparation procedure and analyzed. This blank value must be within three times the standard deviation of the mean blank value determined in the course of the method development. If this is not the case, then the mean blank value must be determined anew as stipulated in Section 5.2.

#### 3.2

##### Operating conditions for atomic absorption spectrometry

The characteristics of the method stated in Section 5 were determined under the following experimental conditions:

**Table 1** Temperature/time program.

Program step	Explanation	Heating rate [°C/s]	Furnace temperature [°C]	Holding time [s]	Argon purging gas
1	Injection	15	90	10	max.
2	Drying	5	105	15	max.
3	Drying	5	120	10	max.
4	Pyrolysis	20	750	5	max.
5	Zero adjustment	0	750	4	min.
6	Atomization/ measurement	1000	1400	3	min.
7	Heating	3000	2600	4	max.

Apparatus:	Atomic absorption spectrometer with equipment for the graphite furnace technique “AAS vario 6 EA” from Analytik Jena, “MPE 50 Mikropipettierereinheit” (micropipetting unit), integrated circulation cooler
Absorption:	283.3 nm
Slit width:	0.8 nm
Lamp current:	3 mA
Background compensation:	Deuterium lamp
Measuring solution:	An aliquot of the sample solution (see Section 3.1) is diluted with three parts of ultrapure water <sup>2)</sup>
Injection volume:	20 µL of the measuring solution Injection is performed by the micropipetting unit

## 4 Evaluation

### 4.1

#### Calibration

In order to obtain the calibration function, the MPE 50 micropipetting unit of the autosampler is used to prepare the following dilutions from Lead Stock Solution 2 (see Section 1.2). Dilution with ultrapure water is performed automatically.

2) As a rule a dilution of 1:4 is sufficient for a measurement free of interference. If the matrix of the measuring solution is still too concentrated or the lead content too high, then it must be further diluted.

**Table 2** Dilutions for the determination of the calibration function

Stock Solution 2 [ $\mu\text{L}$ ]	0	4	8	12	16	20
Ultrapure water [ $\mu\text{L}$ ]	20	16	12	8	4	0
Mass (lead) [ng]	0.0	0.1	0.2	0.3	0.4	0.5
Lead concentration [ $\mu\text{g/L}$ ]		5	10	15	20	25

The measurement is performed with a constant injection volume of 20  $\mu\text{L}$ . The absorbances are determined from the peak heights and recorded versus the concentration. Under the conditions stated here the calibration function is linear.

Based on an air sample volume of 1.2  $\text{m}^3$ , a concentration range from approx. 0.4 to 1.7  $\mu\text{g}/\text{m}^3$  is covered by the dilutions stated in Table 2. The measurement range can be extended by appropriate dilution steps. Spectral interference may occur in some cases when multiple-element lamps are used. Therefore lines and signals must always be critically checked.

## 4.2

### Calculation of the analytical result

The corresponding value for the mass in the measuring solution is obtained from the absorbance in the calibration function, and the lead mass in the sample is then calculated according to equation (1).

$$m = (c - \bar{c}_{\text{Bl}(\text{current})}) \times V \times f \quad (1)$$

The concentration by weight of lead in the air sample in  $\mu\text{g}/\text{m}^3$  is calculated according to equation (2).

$$c_m = \frac{m}{V_{\text{Air}}} \quad (2)$$

where:

$m$	is the mass of lead in the sample in $\mu\text{g}$
$c$	is the lead concentration in the test solution in $\mu\text{g/L}$
$c_{\text{Bl}(\text{current})}$	is the current mean value for the lead concentration in the blank solution in $\mu\text{g/L}$ (see Section 3.1)
$V$	is the volume of the sample solution in L
$f$	is the dilution factor ( $f = 4$ , see Section 3.2)
$c_m$	is the lead concentration by weight in the air sample in $\mu\text{g}/\text{m}^3$
$V_{\text{Air}}$	is the air sample volume in $\text{m}^3$



## 5 Reliability of the method

### 5.1

#### Precision and recovery

Six filters were each spiked with three solutions of different lead concentrations and then subjected to the entire analytical procedure.

Lead Stock Solution 1 (see Section 1.2) served as the starting solution for spiking the filters (see Table 3).

**Table 3** Determination of the relative standard deviation.

Spiked mass of lead [µg]	Concentration in air* [µg/m <sup>3</sup> ]	Relative standard deviation [%]
12	10	2.2
60	50	1.0
120	100	0.8

\* The concentration is calculated on the basis of an air sample volume of 1.2 m<sup>3</sup> (2-hour sampling period at a flow rate of 10 L/min)

As a result of the different chemical composition and the different physical properties of the aerosols in various work areas, it is impossible to make generally a valid statement on the recovery for the entire procedure.

On the basis of the sample preparation described above the analytical recovery is defined as 100% according to DIN EN 13890 (restricted to those metals and compounds that are soluble in the stated system) [3, 4].

The described sample preparation procedure was checked with elemental lead and one of its defined compounds. For this purpose three samples of approx. 2 mg of lead shavings and three sample approx. 0.5 mg of lead(II) oxide were each subjected to the digestion procedure in 80 mL of the standard acid digestion mixture (see Section 3.1). Solutions without any visible particles were obtained in every case. The subsequent analysis resulted in 99% or 98% respectively of the calculated lead concentration.

### 5.2

#### Limit of quantification

During development of the method the limit of quantification was determined according to the blank value method as stipulated in DIN 32645 [5]. For this purpose ten filters that were not used for sampling were subjected to the entire sample preparation procedure, the mean for the blank value ( $\bar{c}_{BI}$ ), which is caused by the

**Table 4** Limit of quantification and blank value.

Limit of quantification [ $X_{LQ}$ ] in the measuring solution	2.2 µg/L
Mean blank value [ $\bar{c}_{BI}$ ] in the measuring solution	0.29 µg/L
Standard deviation [ $s$ ] of the blank values in the measuring solution	0.19 µg/L
Absolute limit of quantification	0.04 µg
Lowest determinable concentration in the air sample*	0.13 µg/m <sup>3</sup>

\* For an air sample volume of 1.2 m<sup>3</sup>, a sample solution of 20 mL and a dilution factor of 4

filters, reagents and vessels used, as well as the corresponding standard deviation were determined and inserted into equation (3) to calculate the limit of quantification.

$$X_{LQ} = \bar{c}_{BI} + 10 \cdot s \quad (3)$$

After subtraction of the mean blank value, the lowest determinable concentration in the air sample is dependent only on ten times the standard deviation (see Table 4).

### 5.3

#### Selectivity

The selectivity of the method depends above all on the wavelength selected and on the absence of spectral interference. In the case of non-spectral interference, such as more complex matrix effects, it is advisable to use the addition procedure. The stated conditions have proved successful in practice [6].

The sampling system described here is not suitable for vaporous lead or vaporous lead compounds.

If compounds that can occur both as vapor and particles in the air at workplaces, the method must be adapted to meet the requirements of simultaneous sampling by modifying the sampling system (e.g. by combining a filter and a liquid collection phase).

## 6

#### Remarks

The analytical procedure for the determination of lead that has been published in this series [7] is now superseded by the present method.

## References

- 1 Hebisch R, Fricke H-H, Hahn J-U, Lahaniatis M, Maschmeier C-P, Mattenklott M (2005) Sampling and determining aerosols and their chemical compounds. In: Parlar H, Greim H (eds.) *The MAK Collection for Occupational Health and Safety, Part III: Air Monitoring Methods*, Vol. 9: 2–40. Wiley-VCH, Weinheim.
- 2 DIN EN 481 (1993) Workplace atmospheres – Size fraction definitions for measurement of airborne particles. Beuth Verlag, Berlin.
- 3 DIN EN 482 (2006) Workplace atmospheres – General requirements for the performance of procedures for the measurement of chemical agents. Beuth Verlag, Berlin. DIN EN 482
- 4 DIN EN 13890 (2010) Workplace atmospheres – Procedures for measuring metalloids in airborne particles – Requirements and methods. Beuth Verlag, Berlin.
- 5 DIN 32645 (2008) Chemical analysis; decision limit, detection limit and determination limit under repeatability conditions; terms, methods, evaluation. Beuth Verlag, Berlin DIN EN 32645
- 6 Doerffel K, Eckschlagler K (1981) *Optimale Strategien in der Analytik*. Verlag Harri Deutsch, Thun, Frankfurt.
- 7 Angerer J, Knecht U (1991) Lead. In: DFG (Deutsche Forschungsgemeinschaft) (Kettrup A, Henschler D eds.) *Analysis of Hazardous Substances in Air*, Volume 1, 177–187. Wiley-VCH, Weinheim.

Author: J.-U. Hahn

